

## Intended use

The Wako NEFA-HR(2) reagent is an *in vitro* enzymatic colorimetric method assay for the quantitative determination of non-esterified fatty acids (NEFA) in serum.

## Summary and explanation of the test

Non-esterified fatty acid (NEFA) in serum binding albumin, is used as an important energy source of peripheral tissues. The amount of NEFA in serum depends on a balance between intake in liver and peripheral tissues, and the release from adipose tissues. Amount of NEFA decreases by physical exercise, increases by starvation, cold, fear or smoking. And then increase or decrease of NEFA is observed in diabetes, hepatic diseases or endocrine diseases.

NEFA had been assayed by organic solvent extraction method, which was complicated to operate. Enzymatic method using Acyl-CoA oxidase (ACOD) has become widespread due to excellent specificity and concise procedure. NEFA-HR(2) is the reagent kit for NEFA assay based on enzymatic method using 3-Methyl-N-Ethyl-N-( $\beta$ -Hydroxyethyl)-Aniline (MEHA) as a violet color agent.



It gives reliable results without interference from ascorbic acid and bilirubin.

## Principle of the method

Non-esterified fatty acid (NEFA) in sample is converted to Acyl-CoA, AMP and pyrophosphoric acid (PPi) by the action of Acyl-CoA synthetase (ACS), under coexistence with coenzyme A (CoA) and adenosine 5-triphosphate disodium salt (ATP). Obtained Acyl-CoA is oxidized and yields 2,3-trans-Enoyl-CoA and hydrogen peroxide by the action of Acyl-CoA oxidase (ACOD). In the presence of peroxidase (POD), the hydrogen peroxide formed yields a blue purple pigment by quantitative oxidation condensation with 3-Methyl-N-Ethyl-N-( $\beta$ -Hydroxyethyl)-Aniline (MEHA) and 4-aminoantipyrene (4-AA).

Non-esterified fatty acids (NEFA) concentration is obtained by measuring absorbance of the blue purple color.

## Reactions



## Physical or chemical indications of instability

The presence of precipitates in the reagents or values of control sera outside the manufacturer's acceptable range may be an indication of reagent's instability.

## Instruments

The reagent is designed to be used on commercially available automated analyzers. Refer to the operating manual for a description of instrument operation and specifications. A validation by the user in practice at the customer's site in the form of measurements of adequate control or patient sera in sufficient number is indispensable.

## Reagents

### Contents and storage conditions

R1 Set:	R1a:	Color A	Store at 2 - 10°C
	R1:	Solvent A	
R2 Set:	R2a:	Color B	Store at 2 - 10°C
	R2:	Solvent B	

### Ingredients

#### R1 Set:

R1a: Color A	(when reconstituted)	
ACS		0.53 U/mL
CoA		0.31 mmol/L
ATP		4.3 mmol/L
4-AA		1.5 mmol/L
AOD		2.6 U/mL
Sodium azide		0.062%
(Color A lyophilized)		(0.8%)

R1: Solvent A		
Phosphate Buffer, pH 7.0		50 mmol/L
Sodium azide		0.05%

#### R2 Set:

R2a: Color B	(when reconstituted)	
ACOD		12 U/mL
POD		14 U/mL

R2: Solvent B		
MEHA		2.4 mmol/L

## Reagent preparation

- R1: Prepare R1 by mixing one bottle of Color A and Solvent A.  
After preparing the R1, store at 2 - 10°C and use within 1 month.
- R2: Prepare R2 by mixing one bottle of Color B and Solvent B.  
After preparing the R2, store at 2 - 10°C and use within 1 month.

## Specimen collection and preparation


Serum can be used as specimen.

Assay samples immediately after collection, because the enzymes such as lipoprotein lipase, phospholipase etc. hydrolyze lipids and form fatty acids. Freeze sample, when a serum is stored. Stability: 2 days at 4°C.<sup>1</sup>

*In vivo* heparin addition causes wrongly increased values. Because of stimulation of the lipoprotein lipase by heparin samples of patients under heparin treatment blood can be used for this determination only after appropriate pre-treatment.<sup>2</sup>

## Standard procedure

Temperature: 37°C (Hitachi®737)



Calibrator: Wako NEFA Standard (Available separately.)

## Calculation of NEFA concentration

Calculate NEFA concentration from the calibration curve which was created from absorbance of calibrator.

Conversion factors:  $\text{mg/dL} = \text{mmol/L} \times 28.2$   
(calculated for oleic acid, MW = 282)  
 $\text{mmol/L (mval/L)} = \text{mEq/L} = \text{mg/dL} \times 0.035$

## Application to the various automatic analyzers

Input the parameters according to the instructions of instruments to perform the measurement. Instrument applications are available upon request.

## Results

The final results are automatically calculated and printed in concentration. The results are given in mEq/L. Always use the same unit for the calibrator.

## Expected values<sup>3</sup>

Men: 0.1 - 0.60 mmol/L (2.8 - 16.9 mg/dL)  
Women: 0.1 - 0.45 mmol/L (2.8 - 12.7 mg/dL)

Since expected values are affected by age, sex, diet, geographical location and other factors, each laboratory should establish its own expected values for this procedure.

Manufactured by:

Wako Chemicals GmbH  
Fuggerstraße 12, D-41468 Neuss  
Telephon(e): +49-2131-311-0  
Facsimile: +49-2131-311-100  
URL: www.wako-chemicals.de

## Performance characteristics

- (1) **Accuracy**  
When a control serum of known concentration is assayed, the assay value falls within the range of  $\pm 15\%$  of the known concentration.
- (2) **Sensitivity**
  - a) When purified water is assayed, the absorbance is not more than 0.140.
  - b) When a standard of given concentration (oleic acid 1 mEq/L) is assayed, the absorbance is 0.100 - 0.380.
- (3) **Precision**  
When a sample is assayed not less than 5 times in a run, CV of absorbance is not more than 1.5%.
- (4) **Measurement range**  
0.01 - 4.00 mEq/L NEFA. (In the case of using the standard procedure)

## Correlation

Specimen	Serum
Correlation coefficient	$r = 0.997$ (n = 50)
Regression equation	$y = 1.013x - 0.043$
y	Wako NEFA-HR(2) (ACS - ACOD method, mEq/L)
x	Wako NEFA C (ACS - ACOD method, mEq/L)

## Interfering substances

- a) Bilirubin gives slightly negative effect on the assay.
- b) Ascorbic acid and hemolysis do not have significant effects on the assay.
- c) Citrate, oxalate, EDTA and sodium fluoride do not have significant influences on the assay when they are used in their usual amounts.

## Warnings and precautions

- For *in vitro* diagnostic use only.
- The usage and application of this test is reserved for professional use only. Please refer to respective national and local regulations and legislation.
- Not to be used internally in humans or animals.
- Do not use the reagents described above for any purpose other than described herein. Performance cannot be guaranteed if the reagents are used in other procedures or for other purposes.
- Operate the instruments according to operator's manuals under appropriate conditions.
- Store the reagents under the specified conditions. Do not use reagents past the expiration date stated on each reagent container label.
- Do not use reagents which were frozen in error. Such reagents may give false results.
- After opening the reagents, it is recommended to use them immediately. When the opened reagents are stored, cap the bottles and keep them under the specified conditions.
- Do not use the containers and other materials in the package for any purposes other than those described herein.
- The vial is stoppered at reduced pressure. Slowly remove the stopper in order not to release the powder in the vial.
- When the reagent for NEFA is used at the same time as reagent for cholesterol and triglyceride, the cholesterol esterase and lipoprotein lipase in the reagent are adsorbed to cuvettes and may interfere the measured value of NEFA.
- Use NEFA Standard for calibration.
- This assay should not be used as the sole determinant for clinical diagnosis.
- If the reagents come in contact with the mouth, eyes or skin, wash off immediately with a large amount of water.
- Be careful not to cut yourself with the aluminum cap when remove it from the vial.
- When discarding the reagents, dispose of them according to local or national regulations. Solvent A contains 3 mg/L potassium ferrocyanide (1 mg/L as cyan).
- All the devices including reagents and reagent bottles contacted with specimen should be considered potentially infectious.
- Sodium azide may react with lead or copper plumbing to form explosive compounds. Even though the reagent contains minute quantity of sodium azide, drains should be flushed well with a large amount of water, when discarding the reagents.
- Color A (R1a) contains components classified as follows according to the European Directive 1999/ 45/ EC.

## Code letter and hazard designation



Xn Harmful

### Risk phrases:

- R 22 Harmful if swallowed  
R 52/53 Harmful to aquatic organisms, may cause long-term adverse effects in the aquatic environment

### Safety phrases:

- S 28 After contact with skin, wash immediately with plenty of water.  
S 45 In case of accident or if you feel unwell, seek medical advice immediately (show the label where possible)  
S 60 This material and its container must be disposed of as hazardous waste.  
S 61 Avoid release to the environment. Refer to special instructions/safety data sheets.

## Quality control

A quality control program is recommended for all clinical laboratories.

## References

1. Rogiers V. Stability of the long chain non-esterified fatty acid pattern in plasma and blood during different storage conditions. Clin Chim Acta. 84, 49 - 54 (1978).
2. Krebs, M. et al., Prevention of *in Vitro* Lipolysis by Tetrahydrolipstatin. Clin. Chem. 46 (7), 950 - 954 (2000).
3. Aufenanger, J. and Kattermann, R. Klinisch-chemische Meßgröße: Freie Fettsäuren (FFS), S. 319 - 320 in Greiling / Greßner: Lehrbuch der Klinischen Chemie und Pathobiochemie, 3. edition, Schattauer (1995).

## Ordering information

Code No.	Product	Package
434-91795	NEFA-HR(2) R1 Set	R1a: 4 x for 50 mL R1: 4 x 50 mL
436-91995	NEFA-HR(2) R2 Set	R2a: 4 x for 25 mL R2: 4 x 25 mL
270-77000	NEFA Standard	CAL: 2 x 10 mL